## **Medical Physicist's ACR DM QC Test Summary**

Facility Name	M	MAP ID-Unit# -					
Address			Room ID				
		Report Date					
				Su	rvey Date		
X-Ray Unit Manufacturer		ı	Model				
Control Panel Serial #	-	<del></del>			stallation Date		
DM Unit Type: Digital radiography (DR)	Compute	Computed radiography (CR) Digital Breast Tomosynthesis (D					
Unit Use: Diagnostic and screening mamme							
	0 . ,	Diagnostic only		Screen			
Survey Type: Mammography equipment evalua	ation (MEE	:) - Full MEE	- Partial	_	Annual survey		
Equipment Tested: DM unit AW monitor	RW	/ monitor View	box	Printer	Other:		
Oversight Level: Medical physicist on-site	Me	dical physicist oversight					
Quality Control Manual Used for Survey and Facility QC:	20	18 ACR Digital Mammogi	raphy QC I	Manual (	with 2D and DBT QC)		
Medical Physicist		Signature					
	Test	Results					
		Pass/Fail*					
Test	2D**	2D Add-on DBT	DBT	CA			
Medical Physicist Tests		•					
Mammography Equipment Evaluation - MQSA Reqs							
ACR DM Phantom Image Quality					Your Phantom		
3. DBT Z Resolution					Results - 2D		
Spatial Resolution					Fiber (≥ 2.0)		
5. DBT Volume Coverage					Speck grp (≥ 3.0)		
6. Automatic Exposure Control System Performance					Mass (≥ 2.0)		
7. Average Glandular Dose					AGD (≤ 3.0 mGy)		
8. Unit Checklist							
9. Computed Radiography (if applicable)					Your Phantom		
10. Acquisition Workstation Monitor QC					Results - DBT		
11. Radiologist Workstation Monitor QC					Fiber (≥ 2.0)		
12. Film Printer QC (if applicable)	<u> </u>				Speck grp (≥ 3.0)		
13. Evaluation of Site's Technologist QC Program					Mass (≥ 2.0)		
14. Evaluation of Display Device Technologist QC Program					AGD (≤ 3.0 mGy)		
15. Manufacturer Calibrations (if applicable)							
16. Collimation Assessment	<u> </u>						
MEE/Troubleshooting - Beam Quality (HVL) Assessment	<del> </del>						
MEE/Troubleshooting - kVp Accuracy and Reproducibility	<del>                                     </del>						
Troubleshooting - Ghost Image Evaluation							
Troubleshooting - Viewbox Luminance  Technologist QC Evaluation  Date	o roviowo	d if after new unit MEE:					
ACR DM Phantom Image Quality	Teviewe	d if after flew difft MLL.					
Computed Radiography Cassette Erasure (if applicable)	+						
Compression Thickness Indicator	†						
Visual Checklist	+						
Acquisition Workstation Monitor QC	1				* "Pass" means all		
Radiologist Workstation Monitor QC					components of test		
7. Film Printer QC (if applicable)	1				passes; "Fail" means		
8. Viewbox Cleanliness (if applicable)	1				any or all components fail; if "CA" checked,		
9. Facility QC Review	1				see Corrective Action		
10. Compression Force	1				Summary		
11. Manufacturer Calibration (if applicable)	1				** or DBT aquistition only		
Optional - Repeat Analysis	1						

## **Medical Physicist's ACR DM QC Test Summary** (cont)

Facility Name	MAP ID-Unit# (00000-00)	
Mfr & Model	Room ID	
	Survey Date	

## **Corrective Action Summary\***

\*Note: This is only a summary page, the Corrective Action Log Form may contain further details.

Required/ Recommended	Time Frame	Description	Utilize Corrective Action Log Form	Date Completed	Initials
	·				
		,			