

OXYGEN "H" CYLINDER CHART—PEDIATRIC

L P M	PSI	2200	1200	600	300
	1.0	4 days 16 hours	2 days 13 hours	1 day 6 hours	15 hours
	3/4	6 days 5 hours	3 days 2 hours	1 day 13 hours	18 hours
	1/2	9 days 8 hours	5 days	2 days 12 hours	1 day 6 hours
	1/4	16 days	8 days 17 hours	4 days 8 hours	2 days 4 hours
	1/8	32 days	17 days 10 hours	8 days 22 hours	4 days 11 hours
	1/16	64 days	34 days 21 hours	17 days 10 hours	8 days 17 hours
	1/32	128 days	69 days 18 hours	35 days 6 hours	17 days 15 hours

L P M	PSI	2200	1200	600	300
	1.0	4 days 16 hours	2 days 13 hours	1 day 6 hours	15 hours
	0.9	4 days 5 hours	2 days 20 hours	1 day 10 hours	17 hours
	0.8	5 days 20 hours	3 days 4 hours	1 day 14 hours	19 hours
	0.7	6 days 16 hours	3 days 15 hours	1 day 19 hours	21 hours 30 minutes
	0.6	7 days 18 hours	4 days 5 hours	2 days 2 hours	1 day 1 hour
	0.5	9 days 8 hours	5 days 2 hours	2 days 7 hours	1 day 3 hours
	0.4	11 days 16 hours	6 days 8 hours	3 days 4 hours	1 day 14 hours
	0.3	15 days 13 hours	8 days 10 hours	4 days 5 hours	2 days 2 hours
	0.2	23 days 8 hours	12 days 16 hours	6 days 8 hours	3 days 4 hours
	0.1	46 days 16 hours	25 days 8 hours	12 days 20 hours	6 days 10 hours

* Call for a tank change when the last cylinder has less than 24 hours of use left.

OXYGEN "E" CYLINDER CHART—PEDIATRIC

L P M	PSI	2200	1200	600	300
	1.0	10 hours	5 hours 30 minutes	2 hours 45 minutes	1 hour 20 minutes
	3/4	13 hours 20 minutes	7 hours 20 minutes	3 hours 40 minutes	1 hour 46 minutes
	1/2	20 hours 24 minutes	11 hours	5 hour 30 minutes	2 hours 40 minutes
	1/4	1 day 16 hours 48 minutes	22 hours	11 hours	5 hours 20 minutes
	1/8	3 days 9 hours 36 minutes	1 day 20 hours	22 hours	10 hours 40 minutes
	1/16	6 days 19 hours 12 minutes	3 days 15 hours	1 day 20 hours	21 hours 30 minutes
	1/32	13 days 14 hours 24 minutes	7 days 20 hours	3 days 15 hours	43 hours

L P M	PSI	2200	1200	600	300
	1.0	10 hours	5 hours 30 minutes	2 hours 45 minutes	1 hour 20 minutes
	0.9	11 hours	6 hours	3 hours	1 hour 30 minutes
	0.8	12 hours 30 minutes	6 hours 50 minutes	3 hours 25 minutes	1 hour 40 minutes
	0.7	14 hours 16 minutes	7 hours 50 minutes	3 hours 45 minutes	1 hour 50 minutes
	0.6	16 hours 40 minutes	9 hours	4 hours 30 minutes	2 hours 15 minutes
	0.5	20 hours	11 hours	5 hours 30 minutes	2 hours 45 minutes
	0.4	25 hours	13 hours 40 minutes	6 hours 50 minutes	3 hours 25 minutes
	0.3	33 hours 20 minutes	18 hours	9 hours	4 hours 30 minutes
	0.2	2 days 2 hours	27 hours 20 minutes	13 hours 40 minutes	6 hours 50 minutes
	0.1	4 days 4 hours	2 days 6 hours	1 day 3 hours	13 hours 30 minutes

* Call for a tank change when the last cylinder has less than 24 hours of use left.

OXYGEN "D" CYLINDER CHART—PEDIATRIC

L P M	PSI	2200	1200	600	300
	1.0	5 hours 30 minutes	3 hours	1 hour 30 minutes	45 minutes
	3/4	7 hours 20 minutes	4 hours	2 hours	1 hour
	1/2	11 hours	6 hours	3 hours	1 hour 30 minutes
	1/4	22 hours	12 hours	6 hours	3 hours
	1/8	44 hours	24 hours	12 hours	6 hours
	1/16	88 hours	48 hours	24 hours	12 hours
	1/32	166 hours	96 hours	48 hours	24 hours

L P M	PSI	2200	1200	600	300
	1.0	5 hours 30 minutes	3 hours	1 hour 30 minutes	45 minutes
	0.9	6 hours 6 minutes	3 hours 19 minutes	1 hour 39 minutes	50 minutes
	0.8	6 hours 52 minutes	3 hours 44 minutes	1 hour 52 minutes	56 minutes
	0.7	7 hours 51 minutes	4 hours 16 minutes	2 hours 8 minutes	64 minutes
	0.6	9 hours 10 minutes	5 hours	2.5 hours	1 hour 15 minutes
	0.5	11 hours	6 hours	3 hours	1 hour 30 minutes
	0.4	13 hours 45 minutes	7 hours 28 minutes	3 hours 43 minutes	1 hour 50 minutes
	0.3	18 hours 20 minutes	10 hours	5 hours	2 hours 30 minutes
	0.2	27 hours 30 minutes	15 hours	7.5 hours	3 hours 45 minutes
0.1	55 hours	30 hours	15 hours	7.5 hours	

* Call for a tank change when the last cylinder has less than 24 hours of use left.

OXYGEN "M" CYLINDER CHART—PEDIATRIC

L P M	PSI	2200	1200	600	300
	1.0	2 days 2 hours	1 day 3 hours	13 hours 30 minutes	6 hours 45 minutes
	3/4	2 days 19 hours	1 day 9 hours	16 hours 30 minutes	8 hours
	1/2	4 days 4 hours 45 minutes	2 days 6 hours	1 day 3 hours	13 hours 30 minutes
	1/4	11 days 45 minutes	3 days 2 hours	1 day 22 hours	23 hours 15 minutes
	1/8	14 days 9 hours	7 days 19 hours	4 days	2 days
	1/16	28 days 18 hours	15 days 13 hours	7 days 20 hours	3 days 11 hours
	1/32	57 days 12 hours	31 days 2 hours	15 days 20 hours	7 days 3 hours

L P M	PSI	2200	1200	600	300
	1.0	2 days 2 hours	1 day 3 hours 45 minutes	13 hours 30 minutes	6 hours 45 minutes
	0.9	1 day 21 hours	1 day 6 hours	15 hours	7 hours 30 minutes
	0.8	2 days 15 hours	1 day 10 hours	17 hours	8 hours 30 minutes
	0.7	3 days	1 day 15 hours	19 hours 30 minutes	9 hours 45 minutes
	0.6	3 days 11 hours	1 day 21 hours	22 hours 30 minutes	11 hours 15 minutes
	0.5	4 days 4 hours	2 days 6 hours	1 day 3 hours	13 hours 30 minutes
	0.4	5 days 6 hours	2 days 20 hours	1 day 10 hours	17 hours
	0.3	6 days 23 hours	3 days 18 hours	1 day 21 hours	22 hours 30 minutes
	0.2	10 days 12 hours	5 days 16 hours	2 days 19 hours	1 day 10 hours
	0.1	21 days	11 days 9 hours	5 days 16 hours	2 days 19 hours

* Call for a tank change when the last cylinder has less than 24 hours of use left.